



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 1136
 CALIBRATION DATE: 13-Apr-21

SBE 9plus PRESSURE CALIBRATION DATA
 10000 psia S/N 127421

DIGIQUARTZ COEFFICIENTS:


C1 = -4.373304e+004
 C2 = 2.440268e-001
 C3 = 1.442500e-002
 D1 = 3.686900e-002
 D2 = 0.000000e+000
 T1 = 3.002769e+001
 T2 = -1.951630e-004
 T3 = 4.244630e-006
 T4 = 2.954280e-009
 T5 = 0.000000e+000

AD590M, AD590B, SLOPE AND OFFSET:

AD590M = 1.28120e-002
 AD590B = -8.82994e+000
 Slope = 1.00000
 Offset = -2.5573 (dbars)

PRESSURE (PSIA)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT TEMPERATURE (°C)	INSTRUMENT PRESSURE (PSIA)	CORRECTED PRESSURE (PSIA)	RESIDUAL (PSIA)
14.668	33312.00	21.3	18.282	14.573	-0.095
2001.344	34058.90	21.3	2005.048	2001.338	-0.006
3988.667	34787.60	21.3	3992.459	3988.748	0.081
5976.059	35499.00	21.4	5979.833	5976.120	0.061
7963.478	36194.10	21.4	7967.044	7963.330	-0.148
9951.196	36874.10	21.4	9954.851	9951.136	-0.060
7963.394	36194.10	21.4	7967.035	7963.321	-0.073
5975.913	35499.00	21.5	5979.818	5976.105	0.192
3988.611	34787.60	21.5	3992.441	3988.730	0.119
2001.286	34058.90	21.5	2005.033	2001.323	0.037
14.670	33312.00	21.6	18.273	14.564	-0.106

Residual (PSIA) = corrected instrument pressure - reference pressure





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 1572
 CALIBRATION DATE: 18-Jun-21

SBE 3 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

g = 4.31355871e-003
 h = 6.36304659e-004
 i = 2.13072570e-005
 j = 1.94856415e-006
 f0 = 1000.0


BATH TEMP (° C)	INSTRUMENT OUTPUT (Hz)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	2789.084	-1.5000	0.00004
1.0000	2950.155	1.0000	-0.00004
4.5000	3186.811	4.5000	-0.00004
8.0000	3436.813	8.0000	-0.00001
11.5000	3700.522	11.5000	0.00005
15.0000	3978.286	15.0001	0.00007
18.5000	4270.437	18.4999	-0.00006
22.0000	4577.331	22.0000	-0.00002
25.5000	4899.274	25.5000	-0.00000
29.0000	5236.572	29.0000	-0.00001
32.5000	5589.526	32.5000	0.00001

f = Instrument Output (Hz)

Temperature ITS-90 (°C) = $1 / \{g + h[\ln(f_0 / f)] + i[\ln^2(f_0 / f)] + j[\ln^3(f_0 / f)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

Date, Offset (mdeg C)
 ● 18-Jun-21 0.00





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 4107
 CALIBRATION DATE: 02-Sep-20

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.01897953e+001
 h = 1.39751243e+000
 i = -6.71617376e-004
 j = 1.22344702e-004

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.70115	0.00000	0.00000
-1.0003	34.8265	2.80530	5.23143	2.80527	-0.00002
0.9997	34.8266	2.97674	5.34731	2.97676	0.00002
14.9997	34.8255	4.27259	6.15242	4.27261	0.00002
18.4997	34.8243	4.61929	6.35040	4.61927	-0.00001
28.9997	34.8199	5.70281	6.93252	5.70279	-0.00002
32.4997	34.8086	6.07477	7.12130	6.07478	0.00001

f = Instrument Output (kHz)


t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = (g + h * f² + i * f³ + j * f⁴) / 10 (1 + δ * t + ε * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

Date, Slope Correction

● 02-Sep-20 1.0000000



**CALIBRATION
 AFTER
 MODIFICATIONS**



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 3553
 CALIBRATION DATE: 08-Jul-21

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS:
 Soc = 0.4305
 Voffset = -0.5047
 Tau20 = 1.28
 A = -4.3491e-003
 B = 1.4728e-004
 C = -1.7851e-006
 E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS
 D1 = 1.92634e-4
 D2 = -4.64803e-2
 H1 = -3.300000e-2
 H2 = 5.00000e+3
 H3 = 1.45000e+3


BATH OXYGEN (ml/l)	BATH TEMPERATURE (° C)	BATH SALINITY (PSU)	INSTRUMENT OUTPUT (volts)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.18	6.00	0.00	0.825	1.18	-0.00
1.18	12.00	0.00	0.880	1.18	-0.00
1.18	2.00	0.00	0.790	1.18	-0.00
1.18	30.00	0.00	1.052	1.19	0.01
1.18	20.00	0.00	0.955	1.18	-0.00
1.18	26.00	0.00	1.012	1.18	0.00
3.95	30.00	0.00	2.326	3.95	0.00
3.96	2.00	0.00	1.462	3.96	-0.00
3.96	6.00	0.00	1.585	3.96	0.00
3.97	12.00	0.00	1.771	3.97	0.00
3.97	26.00	0.00	2.210	3.98	0.01
3.98	20.00	0.00	2.021	3.98	-0.00
6.75	2.00	0.00	2.139	6.75	-0.00
6.78	6.00	0.00	2.354	6.79	0.00
6.83	12.00	0.00	2.684	6.83	-0.00
6.84	30.00	0.00	3.649	6.83	-0.01
6.88	20.00	0.00	3.126	6.87	-0.00
6.90	26.00	0.00	3.460	6.90	0.00

V = instrument output (volts); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen



ECO Chlorophyll Fluorometer Characterization Sheet

Date: 10/20/2021

S/N: FLRTD-7158

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

	Analog Range 1	Analog Range 2	Analog Range 4 (default)	Digital
Dark Counts	0.067	0.039	0.025 V	46 counts
Scale Factor (SF)	6	13	25 $\mu\text{g/l/V}$	0.0076 $\mu\text{g/l/count}$
Maximum Output	4.99	4.99	4.99 V	16380 counts
Resolution	0.5	0.5	0.5 mV	1.0 counts
Ambient temperature during characterization				22.5 °C

Analog Range: 1 (most sensitive, 0–4,000 counts), 2 (midrange, 0–8,000 counts), 4 (entire range, 0–16,000 counts).

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $\text{SF} = x \div (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations *in-situ* is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.sea-birdscientific.com

C-Star Calibration

Date September 23, 2021 S/N# CST-2151 Pathlength 25 cm

	Analog output	Digital output	
V_{dark}	0.000 V	0 counts	
V_{air}	4.802 V	15784 counts	
V_{ref}	4.702 V	15455 counts	
Temperature of calibration water			23.4 °C
Ambient temperature during calibration			23.2 °C

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x , in meters): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{\text{sig}} - V_{\text{dark}}) / (V_{\text{ref}} - V_{\text{dark}})$

To determine beam attenuation coefficient: $c = -1/x * \ln(Tr)$

V_{dark} Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref} .

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.

CALIBRATION CERTIFICATE
PAR Irradiance Meter




Date of Cal: 19-Oct-2020
Part Number: 0046-3097
Job Number: JN34601i
Serial Number: 19-0003-003

55 Central Avenue
West Molesey
Surrey KT8 2QZ
United Kingdom
T +44 (0)20 8481 9000
E sales@chelsea.co.uk
chelsea.co.uk

Calibration Equation

PAR (Ln(uW.cm-2)) = 0.0047143 x mVolts - 2.893023
PAR(uW.cm-2) = PAR(uE.m-2.s-1) / 0.04234
PAR (Ln(uE.m-2.s-1)) = 0.0047143 x mVolts - 6.055046



Calibration
Coeff A (slope): 0.0047143
Coeff B (intercept): -2.893023
RSQ: 0.9981

Lamp Stability
Within cal: 6.32%
Between cal: 6.39%

Multimeter
Manufacturer: FLUKE 8846A
SN: 4549021
Cal Due Date: Cal. Checked

Supply Volts
12.00

Electal Offset (mv)
0.0

Observations:

Sensor Condition: na
Accessories: Sensor Cap

Instrument calibrated at Chelsea Technologies Group as described in section 3 of the PAR sensor handbook.

Name: LJ Gibson

Signed: 

Issue Date: 19 October 2020